

质量检验报告

Certificate of Analysis

产品名称 Product Name: SPARKeasy尿液基因组DNA快速提取试剂盒		
货号 Catalog No: AA1502	批号 Lot No: SMQQG	
规格 Package: 50 T	储存温度 Store at: RT	
生产日期 Date of Manuf: May.27,2021	有效期 Expiry date: May.26,2022	
检验日期 Inspection date: May.27,2021	报告日期 Report date: May.28,2021	
检测项目 Test	检测标准 Specifications	检测结果 Results of analysis
外观 Appearance	内外包装完整, 标签正确 The inner and outer packaging is complete and the label is correct.	合格 Qualified
说明书 Instructions	内容准确 Accurate content.	合格 Qualified
性能检测 performance detection	批间差检测: 分别使用待检试剂盒和对照试剂盒提取 30mL 尿液, DNA 产量比值在 90%≤待检/对照≤110%	102%
	纯度检测: A260/A280 在 1.7-1.9 之间, A260/A230 在 2.0-2.5 之间	合格 Qualified
	完整性检测: 琼脂糖凝胶结果显示单一清晰条带	合格 Qualified
结论 Conclusion	结果符合产品质量要求 The results meet the quality requirements.	

质检员 QC Checker:

潘欢

审核 Manager:

陈雪梅

质检专用章